



REHASENSE

**EU DECLARATION OF CONFORMITY
REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE
COUNCIL OF 05 APRIL 2017 ON MEDICAL DEVICES (MDR)**

Manufacturer:

Rehasense Sp. z o.o.

Sulejowska 45G

97-300 Piotrków Trybunalski, Poland

SRN: PL-MF-000004772

Declare with sole responsibility that product (a light four-wheeled support for the disabled)

Product name: **PIXEL**

Catalog number: **PRaab700cc**

(aa – colour, b- size, cc- accesories)

Intended use: The rollator has been designed as a tool to assist walking for people who have mobility problems.

Basic UDI-DI: 59074678ROL6U

meet requirements of the Regulation 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices and applicable international standards: ISO14971:2019; ISO 20417:2021; ISO 11199-2:2005;

Class of the medical device 1, in accordance with rule 1 (technical aid for disabled person). The product classification was carried out in accordance with the rules at Annex VIII of the Regulation 2017/745.

Manufacturer declares that follows conformity assessments procedure described in art. 52 para. 7 of the Regulation 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices after drawing up the technical documentation set out at Annexes II and III of the Regulation 2017/745.



Rehasense Sp. z o.o.
Prezes Zarządu

Roger Spencer Dutton

02-08-2021/ Piotrków Trybunalski/ CEO Roger Spencer Dutton



REHASENSE

Rehasense Sp. z o. o.

ul. Sulejowska 45g, 97-300 Piotrków Tryb.

NIP 677-237-14-61, REGON 122658133